

Birth date

Operator

**BAYDOSOVA**

Date

**11/09/12 13:07**

Examination time

**00:02:32**

Examination

**M**

Stiffness (KPa)

**7.2**

IQR (KPa)

**0.7**


CS (KPa)

**7.5**

Success rate

**71%**

mm  
20.0  
30.0  
40.0  
50.0  
60.0  
70.0  
80.0



mm  
20.0  
30.0  
40.0  
50.0  
60.0  
70.0  
80.0



- (01) E = 7.5 KPa
- (02) E = 7.3 KPa
- (03) E = 8.1 KPa
- (04) E = 8.1 KPa
- (05) E = 7.1 KPa
- (06) E = 7.2 KPa
- (07) E = 6.6 KPa
- (08) E = 7.2 KPa
- (09) E = 5.7 KPa
- (10) E = 5.7 KPa

*Cooper's artery  
F1 (1-2)  
no metastasis*

Test # 1

Valid 10

Invalid 4

Fibroscan® is a medical device to be used as a diagnostic aid. Measurements should be performed by a certified operator. Results should be interpreted by a liver specialist doctor according to the clinical context and taking into account the number of valid measurements, their dispersion (IQR) and the success rate.